



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122822\
Data File : VD075058.D
Acq On : 28 Dec 2022 10:17
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D12222S.M
Title : SW846 8260
Last Update : Fri Dec 23 05:54:19 2022



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	15649	PASS
75	95	30	60	51.1	40693	PASS
95	95	100	100	100.0	79576	PASS
96	95	5	9	6.4	5099	PASS
173	174	0.00	2	1.4	806	PASS
174	95	50	100	74.4	59187	PASS
175	174	5	9	8.1	4787	PASS
176	174	95	101	96.8	57320	PASS
177	176	5	9	6.8	3922	PASS