

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011023\
 Data File : VD075109.D
 Acq On : 10 Jan 2023 17:07
 Operator : KP/SY
 Sample : 01094-01
 Misc : 6.79G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 STOCKPILE-A

Quant Time: Jan 11 01:51:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 10 03:36:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	117330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	252358	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	234574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	100453	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	75887	64.514	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.020%
35) Dibromofluoromethane	7.810	113	78914	49.713	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.420%
50) Toluene-d8	10.269	98	309330	49.146	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.300%
62) 4-Bromofluorobenzene	12.575	95	104415	51.684	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011023\
Data File : VD075109.D
Acq On : 10 Jan 2023 17:07
Operator : KP/SY
Sample : 01094-01
Misc : 6.79G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
STOCKPILE-A

Quant Time: Jan 11 01:51:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 10 03:36:33 2023
Response via : Initial Calibration

