

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011023\
 Data File : VD075111.D
 Acq On : 10 Jan 2023 18:03
 Operator : KP/SY
 Sample : 01094-03
 Misc : 6.79G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 STOCKPILE-B

Quant Time: Jan 11 01:52:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 10 03:36:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	35120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	76064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	75614	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	34790	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	30898	87.754	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	175.500%	#
35) Dibromofluoromethane	7.804	113	26846	56.109	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	112.220%	
50) Toluene-d8	10.269	98	97690	51.494	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	102.980%	
62) 4-Bromofluorobenzene	12.575	95	35423	58.172	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	116.340%	
Target Compounds						
16) Acetone	4.022	43	2214	26.538	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011023\
Data File : VD075111.D
Acq On : 10 Jan 2023 18:03
Operator : KP/SY
Sample : 01094-03
Misc : 6.79G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
STOCKPILE-B

Quant Time: Jan 11 01:52:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 10 03:36:33 2023
Response via : Initial Calibration

