

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011122\
 Data File : VD071672.D
 Acq On : 11 Jan 2022 13:42
 Operator : VA/SY
 Sample : N1018-06
 Misc : 4.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FM-E3-02-4.5

Quant Time: Jan 12 04:27:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	368181	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	681945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	605213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	247153	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	234912	49.586	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.180%
35) Dibromofluoromethane	7.909	113	218280	48.433	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.860%
50) Toluene-d8	10.338	98	876827	51.974	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.940%
62) 4-Bromofluorobenzene	12.626	95	257038	43.238	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011122\
Data File : VD071672.D
Acq On : 11 Jan 2022 13:42
Operator : VA/SY
Sample : N1018-06
Misc : 4.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FM-E3-02-4.5

Quant Time: Jan 12 04:27:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

