

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
 Data File : VD075115.D
 Acq On : 11 Jan 2023 10:42
 Operator : KP/SY
 Sample : VD0111SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0111SBL01

Quant Time: Jan 12 04:27:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 10 03:36:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	104484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	226651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	204480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	87533	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	63387	60.512	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.020%
35) Dibromofluoromethane	7.804	113	70449	49.414	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.820%
50) Toluene-d8	10.269	98	278615	49.287	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.580%
62) 4-Bromofluorobenzene	12.575	95	88799	48.940	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
Data File : VD075115.D
Acq On : 11 Jan 2023 10:42
Operator : KP/SY
Sample : VD0111SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0111SBL01

Quant Time: Jan 12 04:27:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 10 03:36:33 2023
Response via : Initial Calibration

