

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
 Data File : VD075119.D
 Acq On : 11 Jan 2023 12:34
 Operator : KP/SY
 Sample : 01109-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-011023

Quant Time: Jan 12 04:28:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 10 03:36:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	3486	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	8405	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	7637	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	2965	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	3281	93.880	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	187.760%#
35) Dibromofluoromethane	7.805	113	2449	46.321	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.640%
50) Toluene-d8	10.269	98	9923	47.336	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.581	95	2969	44.125	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
Data File : VD075119.D
Acq On : 11 Jan 2023 12:34
Operator : KP/SY
Sample : 01109-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-011023

Quant Time: Jan 12 04:28:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 10 03:36:33 2023
Response via : Initial Calibration

