

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
 Data File : VD075120.D
 Acq On : 11 Jan 2023 14:03
 Operator : KP/SY
 Sample : 01069-01
 Misc : 4.35G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-7

Quant Time: Jan 12 04:28:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 10 03:36:33 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	93996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	205719	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	159604	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	40865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	52837	56.069	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.140%
35) Dibromofluoromethane	7.805	113	62329	48.167	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.340%
50) Toluene-d8	10.269	98	250825	48.886	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.780%
62) 4-Bromofluorobenzene	12.581	95	57230	34.750	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.500%
Target Compounds						
20) Methylene Chloride	4.799	84	25560	14.556	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
Data File : VD075120.D
Acq On : 11 Jan 2023 14:03
Operator : KP/SY
Sample : 01069-01
Misc : 4.35G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-7

Quant Time: Jan 12 04:28:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 10 03:36:33 2023
Response via : Initial Calibration

