

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
 Data File : VD075121.D
 Acq On : 11 Jan 2023 14:30
 Operator : KP/SY
 Sample : 01109-01RE
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-011023RE

Quant Time: Jan 12 04:28:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 10 03:36:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	729	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	1858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	1542	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.522	152	443	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	790	108.092	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	216.180%#
35) Dibromofluoromethane	7.810	113	752	64.343	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	128.680%
50) Toluene-d8	10.263	98	2526	54.510	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.020%
62) 4-Bromofluorobenzene	12.575	95	785	52.776	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
Data File : VD075121.D
Acq On : 11 Jan 2023 14:30
Operator : KP/SY
Sample : 01109-01RE
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-011023RE

Quant Time: Jan 12 04:28:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 10 03:36:33 2023
Response via : Initial Calibration

