

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
 Data File : VD075122.D
 Acq On : 11 Jan 2023 17:16
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jan 12 04:28:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 10 03:36:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.88
34) 1,4-Difluorobenzene	8.775	114	67	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	66	50.000	ug/l	#-0.01
72) 1,4-Dichlorobenzene-d4	13.516	152	190	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.269	98	199	119.087	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	238.180%#
62) 4-Bromofluorobenzene	12.587	95	275	512.707	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	=	1025.420%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
Data File : VD075122.D
Acq On : 11 Jan 2023 17:16
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050EC

Quant Time: Jan 12 04:28:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 10 03:36:33 2023
Response via : Initial Calibration

