



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011123\
Data File : VD075113.D
Acq On : 11 Jan 2023 08:56
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010923S.M
Title : SW846 8260
Last Update : Tue Jan 10 03:36:33 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	12321	PASS
75	95	30	60	52.1	31315	PASS
95	95	100	100	100.0	60147	PASS
96	95	5	9	6.6	3953	PASS
173	174	0.00	2	1.4	598	PASS
174	95	50	100	71.6	43077	PASS
175	174	5	9	7.6	3262	PASS
176	174	95	101	98.1	42267	PASS
177	176	5	9	7.1	2994	PASS