

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011222\
 Data File : VD071678.D
 Acq On : 12 Jan 2022 10:48
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 13 05:27:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.411	0.380	7.5	99	0.00
3 P	Chloromethane	0.541	0.508	6.1	100	0.00
4 C	Vinyl Chloride	0.622	0.587	5.6#	99	0.00
5 T	Bromomethane	0.467	0.415	11.1	104	0.00
6 T	Chloroethane	0.443	0.430	2.9	102	0.00
7 T	Trichlorofluoromethane	0.881	0.833	5.4	101	0.00
8 T	Diethyl Ether	0.232	0.227	2.2	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.501	2.7	101	0.00
10 T	Methyl Iodide	0.493	0.458	7.1	88	0.00
11 T	Tert butyl alcohol	0.061	0.036	41.0#	108	0.00
12 CM	1,1-Dichloroethene	0.447	0.419	6.3#	97	0.00
13 T	Acrolein	0.063	0.065	-3.2	97	0.00
14 T	Allyl chloride	0.886	0.887	-0.1	105	0.00
15 T	Acrylonitrile	0.122	0.127	-4.1	104	0.00
16 T	Acetone	0.149	0.165	-10.7	109	0.00
17 T	Carbon Disulfide	1.186	1.050	11.5	93	0.00
18 T	Methyl Acetate	0.465	0.443	4.7	103	0.00
19 T	Methyl tert-butyl Ether	1.180	1.209	-2.5	104	0.00
20 T	Methylene Chloride	0.652	0.542	16.9	100	0.00
21 T	trans-1,2-Dichloroethene	0.510	0.497	2.5	100	0.00
22 T	Diisopropyl ether	1.908	2.027	-6.2	105	0.00
23 T	Vinyl Acetate	0.891	0.991	-11.2	106	0.00
24 P	1,1-Dichloroethane	1.089	1.081	0.7	101	-0.01
25 T	2-Butanone	0.175	0.192	-9.7	112	0.00
26 T	2,2-Dichloropropane	0.915	0.922	-0.8	103	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.608	-1.0	101	0.00
28 T	Bromochloromethane	0.531	0.535	-0.8	97	0.00
29 T	Tetrahydrofuran	0.101	0.108	-6.9	107	0.00
30 C	Chloroform	1.114	1.135	-1.9#	104	0.00
31 T	Cyclohexane	0.959	0.855	10.8	97	0.00
32 T	1,1,1-Trichloroethane	0.982	1.002	-2.0	105	0.00
33 S	1,2-Dichloroethane-d4	0.643	0.653	-1.6	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.330	0.337	-2.1	94	0.00
36 T	1,1-Dichloropropene	0.471	0.476	-1.1	102	0.00
37 T	Ethyl Acetate	0.219	0.230	-5.0	103	0.00
38 T	Carbon Tetrachloride	0.477	0.497	-4.2	103	0.00
39 T	Methylcyclohexane	0.509	0.509	0.0	98	0.00
40 TM	Benzene	1.299	1.351	-4.0	103	0.00
41 T	Methacrylonitrile	0.143	0.150	-4.9	126	0.00
42 TM	1,2-Dichloroethane	0.440	0.460	-4.5	104	0.00
43 T	Isopropyl Acetate	0.412	0.450	-9.2	107	0.00
44 TM	Trichloroethene	0.334	0.328	1.8	98	0.00
45 C	1,2-Dichloropropane	0.351	0.372	-6.0#	104	0.00
46 T	Dibromomethane	0.180	0.189	-5.0	103	0.00
47 T	Bromodichloromethane	0.493	0.526	-6.7	104	0.00
48 T	Methyl methacrylate	0.208	0.210	-1.0	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011222\
 Data File : VD071678.D
 Acq On : 12 Jan 2022 10:48
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleID :
 VSTDCCC050

Quant Time: Jan 13 05:27:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	105	0.00
50 S	Toluene-d8	1.237	1.319	-6.6	93	0.00
51 T	4-Methyl-2-Pentanone	0.218	0.242	-11.0	107	0.00
52 CM	Toluene	0.817	0.872	-6.7#	104	0.00
53 T	t-1,3-Dichloropropene	0.433	0.485	-12.0	111	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.559	-9.6	108	0.00
55 T	1,1,2-Trichloroethane	0.236	0.260	-10.2	106	0.00
56 T	Ethyl methacrylate	0.289	0.325	-12.5	106	0.00
57 T	1,3-Dichloropropane	0.435	0.470	-8.0	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.185	-16.4	111	0.00
59 T	2-Hexanone	0.152	0.180	-18.4	113	0.00
60 T	Dibromochloromethane	0.300	0.326	-8.7	105	0.00
61 T	1,2-Dibromoethane	0.225	0.242	-7.6	105	0.00
62 S	4-Bromofluorobenzene	0.436	0.452	-3.7	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.294	0.298	-1.4	103	0.00
65 PM	Chlorobenzene	0.957	0.978	-2.2	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.372	-2.5	107	0.00
67 C	Ethyl Benzene	1.757	1.833	-4.3#	103	0.00
68 T	m/p-Xylenes	0.653	0.686	-5.1	103	0.00
69 T	o-Xylene	0.618	0.649	-5.0	105	0.00
70 T	Styrene	1.060	1.137	-7.3	104	0.00
71 P	Bromoform	0.177	0.191	-7.9	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.593	3.645	-1.4	102	0.00
74 T	N-amyl acetate	0.881	0.919	-4.3	106	0.00
75 P	1,1,2,2-Tetrachloroethane	0.637	0.653	-2.5	108	0.00
76 T	1,2,3-Trichloropropane	0.483	0.473	2.1	112	0.00
77 T	Bromobenzene	0.776	0.782	-0.8	106	0.00
78 T	n-propylbenzene	4.519	4.574	-1.2	103	0.00
79 T	2-Chlorotoluene	2.547	2.548	-0.0	103	0.00
80 T	1,3,5-Trimethylbenzene	3.002	3.055	-1.8	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.173	0.193	-11.6	118	0.00
82 T	4-Chlorotoluene	2.642	2.699	-2.2	103	0.00
83 T	tert-Butylbenzene	2.565	2.595	-1.2	102	0.00
84 T	1,2,4-Trimethylbenzene	2.969	3.055	-2.9	104	0.00
85 T	sec-Butylbenzene	3.929	4.029	-2.5	101	0.00
86 T	p-Isopropyltoluene	3.226	3.330	-3.2	103	0.00
87 T	1,3-Dichlorobenzene	1.602	1.615	-0.8	104	0.00
88 T	1,4-Dichlorobenzene	1.634	1.592	2.6	102	0.00
89 T	n-Butylbenzene	3.231	3.276	-1.4	102	0.00
90 T	Hexachloroethane	0.638	0.656	-2.8	107	0.00
91 T	1,2-Dichlorobenzene	1.373	1.399	-1.9	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.102	0.0	109	0.00
93 T	1,2,4-Trichlorobenzene	0.791	0.829	-4.8	106	0.00
94 T	Hexachlorobutadiene	0.464	0.456	1.7	102	0.00
95 T	Naphthalene	1.401	1.506	-7.5	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011222\
Data File : VD071678.D
Acq On : 12 Jan 2022 10:48
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 13 05:27:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.674	0.693	-2.8	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6