

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
 Data File : VD071702.D
 Acq On : 13 Jan 2022 11:34
 Operator : VA/SY
 Sample : VD0113SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0113SBL01

Quant Time: Jan 13 14:29:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	267872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	467291	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	430632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	188086	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	191385	55.526	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.060%
35) Dibromofluoromethane	7.914	113	163857	53.059	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.120%
50) Toluene-d8	10.332	98	590343	51.067	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.140%
62) 4-Bromofluorobenzene	12.620	95	195304	47.945	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
Data File : VD071702.D
Acq On : 13 Jan 2022 11:34
Operator : VA/SY
Sample : VD0113SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0113SBL01

Quant Time: Jan 13 14:29:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

