

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
 Data File : VD071707.D
 Acq On : 13 Jan 2022 13:53
 Operator : VA/SY
 Sample : N1050-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-236

Quant Time: Jan 13 14:31:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	310839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	564614	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	435749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	134479	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	210121	52.535	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.080%	
35) Dibromofluoromethane	7.914	113	184726	49.506	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.020%	
50) Toluene-d8	10.332	98	664753	47.592	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.180%	
62) 4-Bromofluorobenzene	12.626	95	165328	33.590	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 67.180%	
Target Compounds						
					Qvalue	
16) Acetone	4.168	43	7054	7.639	ug/l	100
17) Carbon Disulfide	4.403	76	5393	1.240	ug/l #	94
19) Methyl tert-butyl Ether	5.479	73	10469	1.427	ug/l #	77
20) Methylene Chloride	4.956	84	26412	5.392	ug/l	98
40) Benzene	8.344	78	77265	5.268	ug/l	97
52) Toluene	10.397	92	9388	1.017	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
Data File : VD071707.D
Acq On : 13 Jan 2022 13:53
Operator : VA/SY
Sample : N1050-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-236

Quant Time: Jan 13 14:31:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

