

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
 Data File : VD071708.D
 Acq On : 13 Jan 2022 14:22
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jan 14 04:05:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	360293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	650556	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	584200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	252361	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	226790	48.920	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.840%
35) Dibromofluoromethane	7.909	113	207903	48.357	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.720%
50) Toluene-d8	10.332	98	828776	51.496	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.000%
62) 4-Bromofluorobenzene	12.626	95	256791	45.281	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
Data File : VD071708.D
Acq On : 13 Jan 2022 14:22
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jan 14 04:05:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

