

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
 Data File : VD071710.D
 Acq On : 13 Jan 2022 15:19
 Operator : VA/SY
 Sample : N1106-14
 Misc : 6.22G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SO-A4-A41-03-6

Quant Time: Jan 14 04:06:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	226848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	421418	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	398607	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	175768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	184012	63.042	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.080%	
35) Dibromofluoromethane	7.908	113	148273	53.239	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.480%	
50) Toluene-d8	10.332	98	539225	51.722	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.440%	
62) 4-Bromofluorobenzene	12.620	95	174854	47.597	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 95.200%	
Target Compounds						
16) Acetone	4.167	43	7178	10.651	ug/l	Qvalue # 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
Data File : VD071710.D
Acq On : 13 Jan 2022 15:19
Operator : VA/SY
Sample : N1106-14
Misc : 6.22G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SO-A4-A41-03-6

Quant Time: Jan 14 04:06:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

