

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
 Data File : VD071713.D
 Acq On : 13 Jan 2022 16:45
 Operator : VA/SY
 Sample : N1097-04
 Misc : 4.90G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1D

Quant Time: Jan 14 04:06:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	143479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	258975	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	252938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	114562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	134997	73.123	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.240%
35) Dibromofluoromethane	7.909	113	96662	56.478	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.960%
50) Toluene-d8	10.332	98	337235	52.638	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.280%
62) 4-Bromofluorobenzene	12.620	95	115336	51.089	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.180%
Target Compounds						Qvalue
16) Acetone	4.162	43	10688	25.075	ug/l	95
17) Carbon Disulfide	4.415	76	2190	1.148	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
Data File : VD071713.D
Acq On : 13 Jan 2022 16:45
Operator : VA/SY
Sample : N1097-04
Misc : 4.90G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1D

Quant Time: Jan 14 04:06:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

