

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
 Data File : VD071716.D
 Acq On : 13 Jan 2022 18:11
 Operator : VA/SY
 Sample : N1112-04
 Misc : 3.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-4D

Quant Time: Jan 14 04:07:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	221938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	406145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	376833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	167991	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	173537	60.768	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.540%
35) Dibromofluoromethane	7.908	113	141133	52.581	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.160%
50) Toluene-d8	10.332	98	523595	52.112	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.220%
62) 4-Bromofluorobenzene	12.626	95	168366	47.555	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.100%
Target Compounds						Qvalue
16) Acetone	4.161	43	8342	12.653	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
Data File : VD071716.D
Acq On : 13 Jan 2022 18:11
Operator : VA/SY
Sample : N1112-04
Misc : 3.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-4D

Quant Time: Jan 14 04:07:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

