

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
 Data File : VD071718.D
 Acq On : 13 Jan 2022 19:09
 Operator : VA/SY
 Sample : N1136-04
 Misc : 5.52G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-6D

Quant Time: Jan 14 04:07:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	26423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	46145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	45443	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	20336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	27661	81.359	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 162.720%	
35) Dibromofluoromethane	7.903	113	18169	59.578	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 119.160%	
50) Toluene-d8	10.332	98	56067	49.114	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.220%	
62) 4-Bromofluorobenzene	12.626	95	19359	48.126	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.260%	
Target Compounds						
16) Acetone	4.168	43	3579	45.595	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011322\
Data File : VD071718.D
Acq On : 13 Jan 2022 19:09
Operator : VA/SY
Sample : N1136-04
Misc : 5.52G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-6D

Quant Time: Jan 14 04:07:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

