



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD011624\
Data File : VD078167.D
Acq On : 16 Jan 2024 08:47
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D0108245.M
Title : SW846 8260
Last Update : Tue Jan 09 01:29:14 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	13306	PASS
75	95	30	60	54.1	33971	PASS
95	95	100	100	100.0	62803	PASS
96	95	5	9	6.5	4056	PASS
173	174	0.00	2	1.8	908	PASS
174	95	50	100	81.4	51120	PASS
175	174	5	9	7.0	3567	PASS
176	174	95	101	96.5	49315	PASS
177	176	5	9	7.1	3481	PASS