

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
 Data File : VD071776.D
 Acq On : 24 Jan 2022 14:25
 Operator : VA/SY
 Sample : N1243-06
 Misc : 4.78G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 AMI-2

Quant Time: Jan 25 04:11:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	363062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	675131	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	598619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	247440	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	229279	49.080	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.160%
35) Dibromofluoromethane	7.908	113	216876	48.608	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.220%
50) Toluene-d8	10.338	98	870690	52.131	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.260%
62) 4-Bromofluorobenzene	12.626	95	255465	43.407	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.820%
Target Compounds						
16) Acetone	4.156	43	10563	9.794	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
Data File : VD071776.D
Acq On : 24 Jan 2022 14:25
Operator : VA/SY
Sample : N1243-06
Misc : 4.78G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
AMI-2

Quant Time: Jan 25 04:11:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

