

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
 Data File : VD071779.D
 Acq On : 24 Jan 2022 15:49
 Operator : VA/SY
 Sample : N1247-02
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SS-1A

Quant Time: Jan 25 04:11:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	338389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	645156	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	586043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	238271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	237297	54.500	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.000%	
35) Dibromofluoromethane	7.914	113	215298	50.496	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.000%	
50) Toluene-d8	10.338	98	828973	51.940	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.880%	
62) 4-Bromofluorobenzene	12.626	95	251821	44.776	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.560%	
Target Compounds						
95) Naphthalene	15.384	128	10576	1.584	ug/l	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
Data File : VD071779.D
Acq On : 24 Jan 2022 15:49
Operator : VA/SY
Sample : N1247-02
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SS-1A

Quant Time: Jan 25 04:11:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

