

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
 Data File : VD071774.D
 Acq On : 24 Jan 2022 12:48
 Operator : VA/SY
 Sample : VD0124SBS01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0124SBS01

Quant Time: Jan 25 04:10:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
 Supervised By :Mahesh Dadoda 01/25/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	274279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	474688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	448082	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	221776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	217590	61.654	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 123.300%			
35) Dibromofluoromethane	7.914	113	195427	62.296	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 124.600%			
50) Toluene-d8	10.338	98	706031	60.123	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 120.240%			
62) 4-Bromofluorobenzene	12.626	95	252553	61.033	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 122.060%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.991	85	38529	18.189	ug/l	95
3) Chloromethane	2.215	50	53887	19.131	ug/l	98
4) Vinyl Chloride	2.350	62	61866	18.141	ug/l	98
5) Bromomethane	2.762	94	45911	19.732	ug/l	97
6) Chloroethane	2.921	64	47783	19.662	ug/l	96
7) Trichlorofluoromethane	3.279	101	96613	19.985	ug/l	100
8) Diethyl Ether	3.720	74	25965	20.421	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.103	101	60124	21.278	ug/l	97
10) Methyl Iodide	4.303	142	42726	15.815	ug/l	98
11) Tert butyl alcohol	5.203	59	34411	167.443	ug/l #	85
12) 1,1-Dichloroethene	4.068	96	46588	19.000	ug/l	90
13) Acrolein	3.920	56	21824	63.357	ug/l	95
14) Allyl chloride	4.726	41	95359	19.613	ug/l	96
15) Acrylonitrile	5.426	53	75447	112.281	ug/l	99
16) Acetone	4.156	43	79613	97.708	ug/l	91
17) Carbon Disulfide	4.409	76	105296	17.420	ug/l	99
18) Methyl Acetate	4.726	43	57115	24.192	ug/l	97
19) Methyl tert-butyl Ether	5.479	73	137051	21.174	ug/l	91
20) Methylene Chloride	4.973	84	67770	20.970	ug/l	88
21) trans-1,2-Dichloroethene	5.467	96	54418	19.434	ug/l	93
22) Diisopropyl ether	6.367	45	228269	21.804	ug/l	97
23) Vinyl Acetate	6.309	43	496908	101.708	ug/l	99
24) 1,1-Dichloroethane	6.262	63	127176	21.282	ug/l	98
25) 2-Butanone	7.209	43	109104	113.332	ug/l	96
26) 2,2-Dichloropropane	7.209	77	107809	21.484	ug/l	100
27) cis-1,2-Dichloroethene	7.214	96	69676	21.082	ug/l	99
28) Bromochloromethane	7.550	49	54443	18.702	ug/l	93
29) Tetrahydrofuran	7.561	42	59462	107.368	ug/l	98
30) Chloroform	7.708	83	138515	22.667	ug/l	96
31) Cyclohexane	7.985	56	96452	19.937	ug/l #	88
32) 1,1,1-Trichloroethane	7.908	97	115661	21.475	ug/l	96
36) 1,1-Dichloropropene	8.114	75	89096	19.910	ug/l	98
37) Ethyl Acetate	7.291	43	42259	20.347	ug/l	97
38) Carbon Tetrachloride	8.097	117	94752	20.908	ug/l	98
39) Methylcyclohexane	9.355	83	86503	17.896	ug/l	99
40) Benzene	8.350	78	252376	20.469	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
 Data File : VD071774.D
 Acq On : 24 Jan 2022 12:48
 Operator : VA/SY
 Sample : VD0124SBS01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0124SBS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
 Supervised By :Mahesh Dadoda 01/25/2022

Quant Time: Jan 25 04:10:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.520	41	27594	22.193	ug/l #	87
42) 1,2-Dichloroethane	8.426	62	89837	21.525	ug/l	99
43) Isopropyl Acetate	8.450	43	85529	21.877	ug/l	99
44) Trichloroethene	9.108	130	65968	20.790	ug/l	91
45) 1,2-Dichloropropane	9.385	63	71685	21.508	ug/l	97
46) Dibromomethane	9.473	93	36511	21.402	ug/l	99
47) Bromodichloromethane	9.655	83	105599	22.546	ug/l	98
48) Methyl methacrylate	9.455	41	38099	19.284	ug/l	94
49) 1,4-Dioxane	9.461	88	7684	449.262	ug/l #	75
51) 4-Methyl-2-Pentanone	10.226	43	228123	110.466	ug/l	98
52) Toluene	10.397	92	167750	21.615	ug/l	100
53) t-1,3-Dichloropropene	10.614	75	88724	21.594	ug/l	98
54) cis-1,3-Dichloropropene	10.085	75	101575	20.967	ug/l	99
55) 1,1,2-Trichloroethane	10.797	97	50255	22.471	ug/l	91
56) Ethyl methacrylate	10.655	69	55986	20.423	ug/l	96
57) 1,3-Dichloropropane	10.944	76	88217	21.345	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.938	63	160310	105.931	ug/l	99
59) 2-Hexanone	10.979	43	153228	106.234	ug/l	100
60) Dibromochloromethane	11.138	129	63851	22.430	ug/l	97
61) 1,2-Dibromoethane	11.238	107	46608	21.782	ug/l	98
64) Tetrachloroethene	10.873	164	53628	20.375	ug/l	95
65) Chlorobenzene	11.667	112	174806	20.387	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.738	131	68026	20.923	ug/l	98
67) Ethyl Benzene	11.738	91	316552	20.109	ug/l	100
68) m/p-Xylenes	11.849	106	234182	40.000	ug/l	98
69) o-Xylene	12.173	106	111152	20.059	ug/l	97
70) Styrene	12.191	104	192907	20.310	ug/l	98
71) Bromoform	12.355	173	35616	22.435	ug/l #	99
73) Isopropylbenzene	12.473	105	309019	19.389	ug/l	98
74) N-amyl acetate	12.285	43	79681	20.382	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.720	83	60473	21.413	ug/l	98
76) 1,2,3-Trichloropropane	12.773	75	40456m	18.878	ug/l	
77) Bromobenzene	12.749	156	69895	20.303	ug/l	99
78) n-propylbenzene	12.814	91	389585	19.437	ug/l	99
79) 2-Chlorotoluene	12.896	91	224381	19.862	ug/l	97
80) 1,3,5-Trimethylbenzene	12.949	105	263579	19.794	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.520	75	16182	21.052	ug/l	90
82) 4-Chlorotoluene	12.996	91	236289	20.161	ug/l	99
83) tert-Butylbenzene	13.214	119	218909	19.239	ug/l	98
84) 1,2,4-Trimethylbenzene	13.261	105	258671	19.643	ug/l	99
85) sec-Butylbenzene	13.391	105	344986	19.795	ug/l	100
86) p-Isopropyltoluene	13.508	119	281676	19.685	ug/l	100
87) 1,3-Dichlorobenzene	13.508	146	144037	20.274	ug/l	97
88) 1,4-Dichlorobenzene	13.585	146	143998	19.867	ug/l	99
89) n-Butylbenzene	13.832	91	282717	19.729	ug/l	98
90) Hexachloroethane	14.102	117	58113	20.527	ug/l	100
91) 1,2-Dichlorobenzene	13.879	146	125127	20.550	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.490	75	10018	22.070	ug/l	94
93) 1,2,4-Trichlorobenzene	15.143	180	70766	20.167	ug/l	99
94) Hexachlorobutadiene	15.255	225	40988	19.925	ug/l	96
95) Naphthalene	15.385	128	123048	19.805	ug/l	99
96) 1,2,3-Trichlorobenzene	15.579	180	60453	20.218	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
Data File : VD071774.D
Acq On : 24 Jan 2022 12:48
Operator : VA/SY
Sample : VD0124SBS01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022

Quant Time: Jan 25 04:10:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 07 15:39:14 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012422\
 Data File : VD071774.D
 Acq On : 24 Jan 2022 12:48
 Operator : VA/SY
 Sample : VD0124SBS01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1


Instrument :
 MSVOA_D
 ClientSampleId :
 VD0124SBS01

Quant Time: Jan 25 04:10:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D010722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 07 15:39:14 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By : Semsettin Yesilyurt 01/25/2022
 Supervised By : Mahesh Dadoda 01/25/2022





Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument : MSVOA_D
ClientSampleId : VD0124SBS01



Tgt Ion:168 Resp: 274279
Ion Ratio Lower Upper
168 100
99 64.4 51.8 77.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



Dichlorodifluoromethane
Concen: 18.189 ug/l
RT: 1.991 min Scan# 12
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion: 85 Resp: 38529
Ion Ratio Lower Upper
85 100
87 34.4 15.9 47.6





#3
Chloromethane
Concen: 19.131 ug/l
RT: 2.215 min Scan# 50
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 50 Resp: 5388
Ion Ratio Lower Upper
50 100
52 30.3 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#4
Vinyl Chloride
Concen: 18.141 ug/l
RT: 2.350 min Scan# 73
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 62 Resp: 61866
Ion Ratio Lower Upper
62 100
64 30.2 23.2 34.8





#5
Bromomethane
Concen: 19.732 ug/l
RT: 2.762 min Scan# 145
Delta R.T. -0.011 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01



Tgt Ion: 94 Resp: 4591
Ion Ratio Lower Upper
94 100
96 88.8 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#6
Chloroethane
Concen: 19.662 ug/l
RT: 2.921 min Scan# 170
Delta R.T. -0.005 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion: 64 Resp: 47783
Ion Ratio Lower Upper
64 100
66 32.4 24.2 36.2





#7
Trichlorofluoromethane
Concen: 19.985 ug/l
RT: 3.279 min Scan# 21
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion:101 Resp: 9661
Ion Ratio Lower Upper
101 100
103 62.2 49.8 74.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#8
Diethyl Ether
Concen: 20.421 ug/l
RT: 3.720 min Scan# 306
Delta R.T. 0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion: 74 Resp: 25965
Ion Ratio Lower Upper
74 100
45 112.9 56.5 169.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.278 ug/l
RT: 4.103 min Scan# 372
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion:101 Resp: 60124
Ion Ratio Lower Upper
101 100
85 45.5 38.7 58.1
151 68.7 56.2 84.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#10
Methyl Iodide
Concen: 15.815 ug/l
RT: 4.303 min Scan# 405
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Tgt Ion:142 Resp: 42726
Ion Ratio Lower Upper
142 100
127 47.8 39.4 59.2
141 16.1 11.9 17.9





#11

Tert butyl alcohol

Concen: 167.443 ug/l

RT: 5.203 min Scan# 51

Delta R.T. -0.005 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Instrument :

MSVOA_D

ClientSampleId :

VD0124SBS01

Tgt Ion: 59 Resp: 3441

Ion Ratio Lower Upper

59 100





57 3.5 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022

Supervised By :Mahesh Dadoda 01/25/2022



#12

1,1-Dichloroethene

Concen: 19.000 ug/l

RT: 4.068 min Scan# 365

Delta R.T. -0.011 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48




Tgt Ion: 96 Resp: 46588


Ion Ratio Lower Upper

96 100

61 188.6 137.8 206.8

98 64.4 48.3 72.5





#13

Acrolein

Concen: 63.357 ug/l

RT: 3.920 min Scan# 341

Delta R.T. -0.006 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Instrument :

MSVOA_D

ClientSampleId :

VD0124SBS01

Tgt Ion: 56 Resp: 21824

Ion Ratio Lower Upper

56 100





55 74.9 56.9 85.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022

Supervised By :Mahesh Dadoda 01/25/2022



#14

Allyl chloride

Concen: 19.613 ug/l

RT: 4.726 min Scan# 477

Delta R.T. 0.006 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48




Tgt Ion: 41 Resp: 95359


Ion Ratio Lower Upper

41 100

39 83.0 63.2 94.8


76 32.9 25.8 38.6





#15
Acrylonitrile
Concen: 112.281 ug/l
RT: 5.426 min Scan# 595
Delta R.T. 0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01



Tgt Ion: 53 Resp: 7544
Ion Ratio Lower Upper
53 100
52 82.3 65.6 98.4
51 37.2 31.2 46.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#16
Acetone
Concen: 97.708 ug/l
RT: 4.156 min Scan# 380
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion: 43 Resp: 79613
Ion Ratio Lower Upper
43 100
58 30.7 21.0 31.6





#17
Carbon Disulfide
Concen: 17.420 ug/l
RT: 4.409 min Scan# 41
Delta R.T. -0.005 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 76 Resp: 105290
Ion Ratio Lower Upper
76 100
78 8.9 7.4 11.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#18
Methyl Acetate
Concen: 24.192 ug/l
RT: 4.726 min Scan# 477
Delta R.T. 0.012 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 43 Resp: 57115
Ion Ratio Lower Upper
43 100
74 19.2 16.3 24.5





#19
Methyl tert-butyl Ether
Concen: 21.174 ug/l
RT: 5.479 min Scan# 604
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01



Tgt Ion: 73 Resp: 13705
Ion Ratio Lower Upper
73 100
57 19.4 19.2 28.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#20
Methylene Chloride
Concen: 20.970 ug/l
RT: 4.973 min Scan# 519
Delta R.T. 0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion: 84 Resp: 67770
Ion Ratio Lower Upper
84 100
49 131.3 118.6 178.0
51 38.9 37.8 56.6
86 55.1 47.9 71.9





#21
trans-1,2-Dichloroethene
Concen: 19.434 ug/l
RT: 5.467 min Scan# 605
Delta R.T. -0.012 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 96 Resp: 54418
Ion Ratio Lower Upper
96 100
61 137.3 119.4 179.0
98 62.5 49.3 73.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#22
Diisopropyl ether
Concen: 21.804 ug/l
RT: 6.367 min Scan# 756
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 45 Resp: 228269
Ion Ratio Lower Upper
45 100
43 57.2 43.0 64.6
87 23.8 19.0 28.4
59 10.1 8.6 13.0





#23
 Vinyl Acetate
 Concen: 101.708 ug/l
 RT: 6.309 min Scan# 746
 Delta R.T. 0.001 min
 Lab File: VD071774.D
 Acq: 24 Jan 2022 12:48




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0124SBS01



Tgt Ion: 43 Resp: 496908
 Ion Ratio Lower Upper
 43 100
 86 8.8 6.7 10.1




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
 Supervised By :Mahesh Dadoda 01/25/2022



#24
 1,1-Dichloroethane
 Concen: 21.282 ug/l
 RT: 6.262 min Scan# 738
 Delta R.T. -0.011 min
 Lab File: VD071774.D
 Acq: 24 Jan 2022 12:48

Tgt Ion: 63 Resp: 127176
 Ion Ratio Lower Upper
 63 100
 98 8.0 3.4 10.1
 100 5.1 2.5 7.5









#25
2-Butanone
Concen: 113.332 ug/l
RT: 7.209 min Scan# 899
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion: 43 Resp: 109104
Ion Ratio Lower Upper
43 100
72 18.5 16.4 24.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#26
2,2-Dichloropropane
Concen: 21.484 ug/l
RT: 7.209 min Scan# 899
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion: 77 Resp: 107809
Ion Ratio Lower Upper
77 100
97 21.3 10.7 32.1





#27
cis-1,2-Dichloroethene
Concen: 21.082 ug/l
RT: 7.214 min Scan# 900
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 96 Resp: 69670
Ion Ratio Lower Upper
96 100
61 162.7 0.0 322.8
98 64.6 0.0 130.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#28
Bromochloromethane
Concen: 18.702 ug/l
RT: 7.550 min Scan# 957
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 49 Resp: 54443
Ion Ratio Lower Upper
49 100
129 0.5 0.0 4.2
130 66.2 48.9 73.3





#29
Tetrahydrofuran
Concen: 107.368 ug/l
RT: 7.561 min Scan# 91
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 42 Resp: 59462
Ion Ratio Lower Upper
42 100
72 34.9 29.6 44.4
71 34.8 27.3 40.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#30
Chloroform
Concen: 22.667 ug/l
RT: 7.708 min Scan# 984
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 83 Resp: 138515
Ion Ratio Lower Upper
83 100
85 61.9 52.4 78.6





#31
Cyclohexane
Concen: 19.937 ug/l
RT: 7.985 min Scan# 1031
Delta R.T. -0.005 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 56 Resp: 96452
Ion Ratio Lower Upper
56 100
69 35.5 21.7 32.5
84 72.8 65.7 98.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#32
1,1,1-Trichloroethane
Concen: 21.475 ug/l
RT: 7.908 min Scan# 1018
Delta R.T. 0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion: 97 Resp: 115661
Ion Ratio Lower Upper
97 100
99 61.3 51.8 77.6
61 47.8 40.5 60.7





#33

1,2-Dichloroethane-d4

Concen: 61.654 ug/l

RT: 8.326 min Scan# 1090

Delta R.T. -0.006 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Instrument :

MSVOA_D

ClientSampleId :

VD0124SBS01

Tgt Ion: 65 Resp: 217590

Ion Ratio Lower Upper

65 100





67 49.5 0.0 101.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022

Supervised By :Mahesh Dadoda 01/25/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. -0.006 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48




Tgt Ion:114 Resp: 474688


Ion Ratio Lower Upper

114 100

63 24.7 0.0 46.2


88 16.0 0.0 32.0





#35
Dibromofluoromethane
Concen: 62.296 ug/l
RT: 7.914 min Scan# 1019
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 113 Resp: 19542
Ion Ratio Lower Upper
113 100
111 101.8 82.2 123.4
192 18.1 14.2 21.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#36
1,1-Dichloropropene
Concen: 19.910 ug/l
RT: 8.114 min Scan# 1053
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 75 Resp: 89096
Ion Ratio Lower Upper
75 100
110 32.9 16.0 48.0
77 28.9 24.5 36.7





#37
Ethyl Acetate
Concen: 20.347 ug/l
RT: 7.291 min Scan# 914
Delta R.T. -0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01



Tgt Ion: 43 Resp: 42259
Ion Ratio Lower Upper
43 100
61 15.0 11.0 16.4
70 8.5 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#38
Carbon Tetrachloride
Concen: 20.908 ug/l
RT: 8.097 min Scan# 1050
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Tgt Ion:117 Resp: 94752
Ion Ratio Lower Upper
117 100
119 89.6 73.4 110.2
121 29.2 23.9 35.9





#39

Methylcyclohexane

Concen: 17.896 ug/l

RT: 9.355 min Scan# 11

Delta R.T. 0.000 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Instrument :

MSVOA_D

ClientSampleId :

VD0124SBS01





Tgt Ion: 83 Resp: 8650

Ion	Ratio	Lower	Upper
83	100		
55	87.8	70.1	105.1
98	50.8	39.2	58.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#40

Benzene

Concen: 20.469 ug/l

RT: 8.350 min Scan# 1093




Delta R.T. 0.001 min


Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Tgt Ion: 78 Resp: 252376


Ion	Ratio	Lower	Upper
78	100		
77	24.0	19.4	29.0





#41
Methacrylonitrile
Concen: 22.193 ug/l
RT: 7.520 min Scan# 91
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 41 Resp: 27594
Ion Ratio Lower Upper
41 100
39 35.4 46.9 70.3
67 61.1 49.3 73.9
52 26.9 23.1 34.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#42
1,2-Dichloroethane
Concen: 21.525 ug/l
RT: 8.426 min Scan# 1106
Delta R.T. 0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 62 Resp: 89837
Ion Ratio Lower Upper
62 100
98 7.5 0.0 15.4





#43
Isopropyl Acetate
Concen: 21.877 ug/l
RT: 8.450 min Scan# 1110
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 43 Resp: 85529
Ion Ratio Lower Upper
43 100
61 27.0 21.9 32.9
87 11.5 8.9 13.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#44
Trichloroethene
Concen: 20.790 ug/l
RT: 9.108 min Scan# 1222
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion:130 Resp: 65968
Ion Ratio Lower Upper
130 100
95 98.3 0.0 214.4





#45
1,2-Dichloropropane
Concen: 21.508 ug/l
RT: 9.385 min Scan# 1269
Delta R.T. -0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 63 Resp: 71685
Ion Ratio Lower Upper
63 100
65 29.1 24.7 37.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#46
Dibromomethane
Concen: 21.402 ug/l
RT: 9.473 min Scan# 1284
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 93 Resp: 36511
Ion Ratio Lower Upper
93 100
95 81.0 65.1 97.7
174 84.5 68.1 102.1





#47
Bromodichloromethane
Concen: 22.546 ug/l
RT: 9.655 min Scan# 1315
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 83 Resp: 105599
Ion Ratio Lower Upper
83 100
85 62.0 50.6 75.8
127 6.2 6.1 9.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#48
Methyl methacrylate
Concen: 19.284 ug/l
RT: 9.455 min Scan# 1281
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 41 Resp: 38099
Ion Ratio Lower Upper
41 100
69 80.7 59.5 89.3
39 72.9 56.0 84.0





#49
1,4-Dioxane
Concen: 449.262 ug/l
RT: 9.461 min Scan# 1282
Delta R.T. 0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01



Tgt Ion: 88 Resp: 7684
Ion Ratio Lower Upper
88 100
43 53.2 34.6 52.0
58 103.2 62.0 93.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#50
Toluene-d8
Concen: 60.123 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion: 98 Resp: 706031
Ion Ratio Lower Upper
98 100
100 65.1 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 110.466 ug/l
RT: 10.226 min Scan# 1412
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 43 Resp: 22812
Ion Ratio Lower Upper
43 100
58 33.9 28.2 42.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#52
Toluene
Concen: 21.615 ug/l
RT: 10.397 min Scan# 1441
Delta R.T. -0.005 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion: 92 Resp: 167750
Ion Ratio Lower Upper
92 100
91 172.3 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 21.594 ug/l
 RT: 10.614 min Scan# 1478
 Delta R.T. 0.000 min
 Lab File: VD071774.D
 Acq: 24 Jan 2022 12:48

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0124SBS01

Tgt Ion: 75 Resp: 88724
 Ion Ratio Lower Upper
 75 100
 77 31.0 24.2 36.2




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
 Supervised By :Mahesh Dadoda 01/25/2022



#54
 cis-1,3-Dichloropropene
 Concen: 20.967 ug/l
 RT: 10.085 min Scan# 1388
 Delta R.T. -0.000 min
 Lab File: VD071774.D
 Acq: 24 Jan 2022 12:48


Tgt Ion: 75 Resp: 101575
 Ion Ratio Lower Upper
 75 100
 77 31.0 25.0 37.6
 39 63.5 49.8 74.8





#55
1,1,2-Trichloroethane
Concen: 22.471 ug/l
RT: 10.797 min Scan# 1509
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01






Tgt Ion: 97 Resp: 5025


Ion	Ratio	Lower	Upper
97	100		
83	94.5	68.1	102.1
85	52.1	46.8	70.2
99	66.6	48.6	73.0

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022






#56
Ethyl methacrylate
Concen: 20.423 ug/l
RT: 10.655 min Scan# 1485
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion: 69 Resp: 55986

Ion	Ratio	Lower	Upper
69	100		
41	86.6	65.8	98.6
39	46.3	38.7	58.1





#57

1,3-Dichloropropane

Concen: 21.345 ug/l

RT: 10.944 min Scan# 1533

Delta R.T. 0.001 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Instrument :

MSVOA_D

ClientSampleId :

VD0124SBS01

Tgt Ion: 76 Resp: 8821

Ion Ratio Lower Upper

76 100





78 32.0 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022

Supervised By :Mahesh Dadoda 01/25/2022



#58

2-Chloroethyl Vinyl ether

Concen: 105.931 ug/l

RT: 9.938 min Scan# 1363

Delta R.T. 0.001 min

Lab File: VD071774.D




Acq: 24 Jan 2022 12:48


Tgt Ion: 63 Resp: 160310

Ion Ratio Lower Upper

63 100


106 26.6 20.7 31.1





#59
2-Hexanone
Concen: 106.234 ug/l
RT: 10.979 min Scan# 1540
Delta R.T. -0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 43 Resp: 153228
Ion Ratio Lower Upper
43 100
58 48.2 24.3 72.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#60
Dibromochloromethane
Concen: 22.430 ug/l
RT: 11.138 min Scan# 1567
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion:129 Resp: 63851
Ion Ratio Lower Upper
129 100
127 75.5 39.1 117.3





#61
1,2-Dibromoethane
Concen: 21.782 ug/l
RT: 11.238 min Scan# 11
Delta R.T. -0.005 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion:107 Resp: 46608
Ion Ratio Lower Upper
107 100
109 93.7 76.6 114.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#62
4-Bromofluorobenzene
Concen: 61.033 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 95 Resp: 252553
Ion Ratio Lower Upper
95 100
174 75.1 0.0 150.0
176 72.7 0.0 146.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1652
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion:117 Resp: 448082
Ion Ratio Lower Upper
117 100
82 57.1 46.9 70.3
119 31.5 26.6 39.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#64
Tetrachloroethene
Concen: 20.375 ug/l
RT: 10.873 min Scan# 1522
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion:164 Resp: 53628
Ion Ratio Lower Upper
164 100
166 128.1 100.2 150.2
129 98.0 82.0 123.0
131 101.8 73.7 110.5





#65
Chlorobenzene
Concen: 20.387 ug/l
RT: 11.667 min Scan# 1657
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion:112 Resp: 174800
Ion Ratio Lower Upper
112 100
114 30.5 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 20.923 ug/l
RT: 11.738 min Scan# 1669
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion:131 Resp: 68026
Ion Ratio Lower Upper
131 100
133 93.8 48.1 144.3
119 66.6 33.8 101.4









#67
Ethyl Benzene
Concen: 20.109 ug/l
RT: 11.738 min Scan# 1669
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion: 91 Resp: 316552
Ion Ratio Lower Upper
91 100
106 29.3 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#68
m/p-Xylenes
Concen: 40.000 ug/l
RT: 11.849 min Scan# 1688
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Tgt Ion:106 Resp: 234182
Ion Ratio Lower Upper
106 100
91 213.2 168.6 252.8









#69
o-Xylene
Concen: 20.059 ug/l
RT: 12.173 min Scan# 111152
Delta R.T. -0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion:106 Resp: 111152
Ion Ratio Lower Upper
106 100
91 217.6 110.9 332.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#70
Styrene
Concen: 20.310 ug/l
RT: 12.191 min Scan# 1746
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Tgt Ion:104 Resp: 192907
Ion Ratio Lower Upper
104 100
78 54.2 42.8 64.2
103 56.5 43.6 65.4









#71
Bromoform
Concen: 22.435 ug/l
RT: 12.355 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion:173 Resp: 35610
Ion Ratio Lower Upper
173 100
175 47.8 24.4 73.2
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion:152 Resp: 221776
Ion Ratio Lower Upper
152 100
115 60.7 31.5 94.5
150 164.0 0.0 348.0





#73
Isopropylbenzene
Concen: 19.389 ug/l
RT: 12.473 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion:105 Resp: 309019
Ion Ratio Lower Upper
105 100
120 25.8 12.5 37.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#74
N-ethyl acetate
Concen: 20.382 ug/l
RT: 12.285 min Scan# 1762
Delta R.T. 0.007 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion: 43 Resp: 79681
Ion Ratio Lower Upper
43 100
70 37.7 29.5 44.3
55 22.9 19.1 28.7
61 20.3 17.6 26.4





#75
1,1,2,2-Tetrachloroethane
Concen: 21.413 ug/l
RT: 12.720 min Scan# 1836
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 83 Resp: 60473
Ion Ratio Lower Upper
83 100
131 9.5 5.0 14.9
85 62.4 32.2 96.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#76
1,2,3-Trichloropropane
Concen: 18.878 ug/l m
RT: 12.773 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion: 75 Resp: 40456
Ion Ratio Lower Upper
75 100
77 349.2 0.0 0.0#









#77
Bromobenzene
Concen: 20.303 ug/l
RT: 12.749 min Scan# 1841
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion:156 Resp: 6989
Ion Ratio Lower Upper
156 100
77 202.1 101.4 304.2
158 94.2 46.8 140.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#78
n-propylbenzene
Concen: 19.437 ug/l
RT: 12.814 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Tgt Ion: 91 Resp: 389585
Ion Ratio Lower Upper
91 100
120 22.1 10.9 32.7









#79
2-Chlorotoluene
Concen: 19.862 ug/l
RT: 12.896 min Scan# 1866
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion: 91 Resp: 22438
Ion Ratio Lower Upper
91 100
126 31.8 16.7 50.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#80
1,3,5-Trimethylbenzene
Concen: 19.794 ug/l
RT: 12.949 min Scan# 1875
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion:105 Resp: 263579
Ion Ratio Lower Upper
105 100
120 48.6 24.2 72.6





#81
trans-1,4-Dichloro-2-butene
Concen: 21.052 ug/l
RT: 12.520 min Scan# 1802
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion: 75 Resp: 1618
Ion Ratio Lower Upper
75 100
53 104.7 95.4 143.0
89 48.0 37.2 55.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#82
4-Chlorotoluene
Concen: 20.161 ug/l
RT: 12.996 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion: 91 Resp: 236289
Ion Ratio Lower Upper
91 100
126 31.5 16.1 48.3









#83
tert-Butylbenzene
Concen: 19.239 ug/l
RT: 13.214 min Scan# 1920
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion:119 Resp: 218909
Ion Ratio Lower Upper
119 100
91 70.0 35.6 106.8
134 28.4 13.5 40.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#84
1,2,4-Trimethylbenzene
Concen: 19.643 ug/l
RT: 13.261 min Scan# 1928
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion:105 Resp: 258671
Ion Ratio Lower Upper
105 100
120 44.1 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 19.795 ug/l
 RT: 13.391 min Scan# 1950
 Delta R.T. 0.001 min
 Lab File: VD071774.D
 Acq: 24 Jan 2022 12:48




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0124SBS01




Tgt Ion:105 Resp: 344980
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.3

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
 Supervised By :Mahesh Dadoda 01/25/2022




#86
 p-Isopropyltoluene
 Concen: 19.685 ug/l
 RT: 13.508 min Scan# 1970
 Delta R.T. 0.000 min
 Lab File: VD071774.D
 Acq: 24 Jan 2022 12:48



Tgt Ion:119 Resp: 281676
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.7 38.1
 91 24.7 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 20.274 ug/l

RT: 13.508 min Scan# 1970

Delta R.T. 0.000 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Instrument :

MSVOA_D

ClientSampleId :





VD0124SBS01

Tgt Ion:	146	Resp:	14403
Ion Ratio	Lower	Upper	
146	100		
111	42.2	20.8	62.5
148	65.5	31.3	93.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#88

1,4-Dichlorobenzene

Concen: 19.867 ug/l




RT: 13.585 min Scan# 1983


Delta R.T. 0.001 min

Lab File: VD071774.D

Acq: 24 Jan 2022 12:48

Tgt Ion:	146	Resp:	143998
Ion Ratio	Lower	Upper	
146	100		
111	42.0	20.7	62.1
148	63.7	32.3	96.9









#89
n-Butylbenzene
Concen: 19.729 ug/l
RT: 13.832 min Scan# 2025
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion: 91 Resp: 28271
Ion Ratio Lower Upper
91 100
92 51.5 26.6 79.8
134 23.1 12.1 36.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#90
Hexachloroethane
Concen: 20.527 ug/l
RT: 14.102 min Scan# 2071
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Tgt Ion:117 Resp: 58113
Ion Ratio Lower Upper
117 100
201 64.1 32.1 96.5









#91
1,2-Dichlorobenzene
Concen: 20.550 ug/l
RT: 13.879 min Scan# 2033
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48

Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01

Tgt Ion:146 Resp: 12512
Ion Ratio Lower Upper
146 100
111 42.6 21.8 65.3
148 65.5 31.3 93.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 22.070 ug/l
RT: 14.490 min Scan# 2137
Delta R.T. 0.000 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48


Tgt Ion: 75 Resp: 10018
Ion Ratio Lower Upper
75 100
155 67.6 35.4 106.1
157 96.6 44.9 134.5





#93
1,2,4-Trichlorobenzene
Concen: 20.167 ug/l
RT: 15.143 min Scan# 21
Delta R.T. -0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion:180 Resp: 70760
Ion Ratio Lower Upper
180 100
182 94.5 47.5 142.5
145 33.5 16.4 49.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022




#94
Hexachlorobutadiene
Concen: 19.925 ug/l
RT: 15.255 min Scan# 2267
Delta R.T. 0.006 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Tgt Ion:225 Resp: 40988
Ion Ratio Lower Upper
225 100
223 64.9 31.3 93.9
227 66.8 31.4 94.3





#95
Naphthalene
Concen: 19.805 ug/l
RT: 15.385 min Scan# 2111
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48




Instrument :
MSVOA_D
ClientSampleId :
VD0124SBS01




Tgt Ion:128 Resp: 123048
Ion Ratio Lower Upper
128 100
127 13.1 10.9 16.3
129 11.5 9.1 13.7

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/25/2022
Supervised By :Mahesh Dadoda 01/25/2022



#96
1,2,3-Trichlorobenzene
Concen: 20.218 ug/l
RT: 15.579 min Scan# 2322
Delta R.T. 0.001 min
Lab File: VD071774.D
Acq: 24 Jan 2022 12:48



Tgt Ion:180 Resp: 60453
Ion Ratio Lower Upper
180 100
182 99.5 47.0 141.2
145 35.7 16.4 49.2

