

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012424\
 Data File : VD078226.D
 Acq On : 24 Jan 2024 16:38
 Operator : RP/MD
 Sample : P1237-02
 Misc : 5.24G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2-19.5-20.0

Quant Time: Jan 25 00:28:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 23 05:43:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	124978	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	270353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	270070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	113605	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	62940	44.114	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.220%
35) Dibromofluoromethane	7.811	113	81268	45.674	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.340%
50) Toluene-d8	10.269	98	353542	50.181	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.360%
62) 4-Bromofluorobenzene	12.575	95	105655	51.712	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012424\
Data File : VD078226.D
Acq On : 24 Jan 2024 16:38
Operator : RP/MD
Sample : P1237-02
Misc : 5.24G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2-19.5-20.0

Quant Time: Jan 25 00:28:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 23 05:43:07 2024
Response via : Initial Calibration

