



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012424\
Data File : VD078216.D
Acq On : 24 Jan 2024 10:26
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012224S.M
Title : SW846 8260
Last Update : Tue Jan 23 05:43:07 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1858

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	10050	PASS
75	95	30	60	53.6	26552	PASS
95	95	100	100	100.0	49509	PASS
96	95	5	9	6.4	3160	PASS
173	174	0.00	2	1.1	448	PASS
174	95	50	100	79.5	39379	PASS
175	174	5	9	8.2	3223	PASS
176	174	95	101	97.5	38403	PASS
177	176	5	9	6.6	2529	PASS