

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012721\
 Data File : VD068202.D
 Acq On : 27 Jan 2021 17:08
 Operator : VA/SY
 Sample : M1251-02
 Misc : 6.76G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 400026

Quant Time: Jan 27 23:58:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	500990	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	892449	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	780570	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	311457	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	296353	61.84	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.68%
35) Dibromofluoromethane	7.920	113	271211	54.97	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.94%
50) Toluene-d8	10.344	98	1103158	55.90	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.80%
62) 4-Bromofluorobenzene	12.632	95	344784	48.51	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.02%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012721\
Data File : VD068202.D
Acq On : 27 Jan 2021 17:08
Operator : VA/SY
Sample : M1251-02
Misc : 6.76G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
400026

Quant Time: Jan 27 23:58:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

