

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012721\
 Data File : VD068197.D
 Acq On : 27 Jan 2021 12:06
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 23:55:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	65	0.00
2 T	Dichlorodifluoromethane	0.594	0.574	3.4	65	0.00
3 P	Chloromethane	0.640	0.649	-1.4	69	0.00
4 C	Vinyl Chloride	0.711	0.769	-8.2#	70	0.00
5 T	Bromomethane	0.488	0.500	-2.5	69	0.00
6 T	Chloroethane	0.435	0.479	-10.1	71	0.00
7 T	Trichlorofluoromethane	0.934	0.939	-0.5	66	0.00
8 T	Diethyl Ether	0.275	0.269	2.2	64	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.561	-2.9	67	0.00
10 T	Methyl Iodide	0.527	0.617	-17.1	65	0.00
11 T	Tert butyl alcohol	0.065	0.035	46.2#	53	0.00
12 CM	1,1-Dichloroethene	0.548	0.558	-1.8#	66	0.00
13 T	Acrolein	0.038	0.036	5.3	67	0.00
14 T	Allyl chloride	0.797	0.792	0.6	63	0.00
15 T	Acrylonitrile	0.109	0.115	-5.5	66	0.00
16 T	Acetone	0.085	0.104	-22.4	77	0.00
17 T	Carbon Disulfide	1.764	1.730	1.9	63	0.00
18 T	Methyl Acetate	0.222	0.245	-10.4	71	0.00
19 T	Methyl tert-butyl Ether	1.266	1.276	-0.8	64	0.00
20 T	Methylene Chloride	0.684	0.665	2.8	71	0.00
21 T	trans-1,2-Dichloroethene	0.629	0.632	-0.5	65	0.00
22 T	Diisopropyl ether	1.608	1.613	-0.3	64	0.00
23 T	Vinyl Acetate	0.857	0.882	-2.9	64	0.00
24 P	1,1-Dichloroethane	1.042	1.065	-2.2	65	0.00
25 T	2-Butanone	0.129	0.130	-0.8	65	0.00
26 T	2,2-Dichloropropane	0.986	0.986	0.0	65	0.00
27 T	cis-1,2-Dichloroethene	0.683	0.697	-2.0	65	0.00
28 T	Bromochloromethane	0.355	0.344	3.1	68	0.00
29 T	Tetrahydrofuran	0.083	0.085	-2.4	65	0.00
30 C	Chloroform	1.053	1.072	-1.8#	66	0.00
31 T	Cyclohexane	1.036	1.005	3.0	65	0.00
32 T	1,1,1-Trichloroethane	0.949	0.965	-1.7	65	0.00
33 S	1,2-Dichloroethane-d4	0.478	0.438	8.4	63	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	65	0.00
35 S	Dibromofluoromethane	0.276	0.266	3.6	64	0.00
36 T	1,1-Dichloropropene	0.504	0.531	-5.4	68	0.00
37 T	Ethyl Acetate	0.178	0.185	-3.9	66	0.00
38 T	Carbon Tetrachloride	0.472	0.497	-5.3	67	0.00
39 T	Methylcyclohexane	0.659	0.677	-2.7	65	0.00
40 TM	Benzene	1.410	1.450	-2.8	65	0.00
41 T	Methacrylonitrile	0.108	0.092	14.8	51	0.00
42 TM	1,2-Dichloroethane	0.377	0.390	-3.4	65	0.00
43 T	Isopropyl Acetate	0.372	0.372	0.0	63	0.00
44 TM	Trichloroethene	0.387	0.398	-2.8	65	0.00
45 C	1,2-Dichloropropane	0.345	0.350	-1.4#	64	0.00
46 T	Dibromomethane	0.175	0.184	-5.1	66	0.00
47 T	Bromodichloromethane	0.472	0.483	-2.3	64	0.00
48 T	Methyl methacrylate	0.185	0.188	-1.6	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012721\
 Data File : VD068197.D
 Acq On : 27 Jan 2021 12:06
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 23:55:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 13:19:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	65	0.00
50 S	Toluene-d8	1.106	0.986	10.8	63	0.00
51 T	4-Methyl-2-Pentanone	0.173	0.179	-3.5	65	0.00
52 CM	Toluene	0.919	0.925	-0.7#	64	0.00
53 T	t-1,3-Dichloropropene	0.478	0.479	-0.2	63	0.00
54 T	cis-1,3-Dichloropropene	0.569	0.569	0.0	64	0.00
55 T	1,1,2-Trichloroethane	0.248	0.258	-4.0	66	0.00
56 T	Ethyl methacrylate	0.330	0.333	-0.9	63	0.00
57 T	1,3-Dichloropropane	0.437	0.450	-3.0	65	0.00
58 T	2-Chloroethyl Vinyl ether	0.152	0.162	-6.6	72	0.00
59 T	2-Hexanone	0.120	0.127	-5.8	67	0.00
60 T	Dibromochloromethane	0.311	0.315	-1.3	64	0.00
61 T	1,2-Dibromoethane	0.245	0.249	-1.6	64	0.00
62 S	4-Bromofluorobenzene	0.398	0.354	11.1	61	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	64	0.00
64 T	Tetrachloroethene	0.352	0.356	-1.1	65	0.00
65 PM	Chlorobenzene	1.062	1.086	-2.3	64	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.386	-3.2	64	0.00
67 C	Ethyl Benzene	1.967	2.054	-4.4#	66	0.00
68 T	m/p-Xylenes	0.743	0.770	-3.6	65	0.00
69 T	o-Xylene	0.690	0.710	-2.9	65	0.00
70 T	Styrene	1.177	1.210	-2.8	64	0.00
71 P	Bromoform	0.189	0.194	-2.6	64	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
73 T	Isopropylbenzene	4.160	4.360	-4.8	65	0.00
74 T	N-amyl acetate	0.865	0.873	-0.9	62	0.00
75 P	1,1,2,2-Tetrachloroethane	0.662	0.709	-7.1	66	0.00
76 T	1,2,3-Trichloropropane	0.483	0.446	7.7	52	0.00
77 T	Bromobenzene	0.892	0.929	-4.1	64	0.00
78 T	n-propylbenzene	4.922	5.201	-5.7	66	0.00
79 T	2-Chlorotoluene	2.764	2.875	-4.0	65	0.00
80 T	1,3,5-Trimethylbenzene	3.465	3.606	-4.1	65	0.00
81 T	trans-1,4-Dichloro-2-butene	0.239	0.242	-1.3	63	0.00
82 T	4-Chlorotoluene	2.874	2.979	-3.7	65	0.00
83 T	tert-Butylbenzene	2.936	3.051	-3.9	64	0.00
84 T	1,2,4-Trimethylbenzene	3.439	3.553	-3.3	65	0.00
85 T	sec-Butylbenzene	4.107	4.354	-6.0	66	0.00
86 T	p-Isopropyltoluene	3.736	3.984	-6.6	67	0.00
87 T	1,3-Dichlorobenzene	1.715	1.780	-3.8	65	0.00
88 T	1,4-Dichlorobenzene	1.690	1.756	-3.9	65	0.00
89 T	n-Butylbenzene	3.585	3.827	-6.8	66	0.00
90 T	Hexachloroethane	0.675	0.710	-5.2	64	0.00
91 T	1,2-Dichlorobenzene	1.469	1.539	-4.8	64	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.108	0.110	-1.9	63	0.00
93 T	1,2,4-Trichlorobenzene	1.007	1.055	-4.8	64	0.00
94 T	Hexachlorobutadiene	0.558	0.590	-5.7	66	0.00
95 T	Naphthalene	1.888	1.994	-5.6	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012721\
Data File : VD068197.D
Acq On : 27 Jan 2021 12:06
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jan 27 23:55:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 13:19:08 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.883	-4.0	64	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6