



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012723\
Data File : VD075219.D
Acq On : 27 Jan 2023 11:34
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011823S.M
Title : SW846 8260
Last Update : Thu Jan 19 01:32:27 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	8128	PASS
75	95	30	60	54.3	20565	PASS
95	95	100	100	100.0	37861	PASS
96	95	5	9	6.6	2492	PASS
173	174	0.00	2	1.8	548	PASS
174	95	50	100	79.5	30096	PASS
175	174	5	9	7.7	2331	PASS
176	174	95	101	96.5	29037	PASS
177	176	5	9	7.5	2188	PASS