

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012822\
 Data File : VD071829.D
 Acq On : 28 Jan 2022 17:39
 Operator : VA/SY
 Sample : N1310-01RE
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SOIL-PILERE

Quant Time: Jan 29 04:52:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	134165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	242585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	247417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	108605	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	135331	79.805	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	159.600%
35) Dibromofluoromethane	7.908	113	98560	61.809	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	123.620%
50) Toluene-d8	10.338	98	314183	52.758	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.520%
62) 4-Bromofluorobenzene	12.626	95	106308	51.118	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD012822\
Data File : VD071829.D
Acq On : 28 Jan 2022 17:39
Operator : VA/SY
Sample : N1310-01RE
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SOIL-PILERE

Quant Time: Jan 29 04:52:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

