

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013124\
 Data File : VD078291.D
 Acq On : 31 Jan 2024 14:06
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Feb 01 02:59:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D013024S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 31 00:23:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	1500	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	6432	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	15880	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	10060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	19201	1034.033	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 2068.060%#	
35) Dibromofluoromethane	7.805	113	7425	162.149	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 324.300%#	
50) Toluene-d8	10.275	98	9665	55.749	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 111.500%	
62) 4-Bromofluorobenzene	12.575	95	8869	170.825	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 341.640%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD013124\
Data File : VD078291.D
Acq On : 31 Jan 2024 14:06
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Feb 01 02:59:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D013024S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 31 00:23:33 2024
Response via : Initial Calibration

