

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020222\
 Data File : VD071883.D
 Acq On : 02 Feb 2022 17:02
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 01:08:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.593	0.531	10.5	87	0.00
3 P	Chloromethane	0.710	0.660	7.0	91	0.00
4 C	Vinyl Chloride	0.743	0.701	5.7#	89	0.00
5 T	Bromomethane	0.490	0.417	14.9	84	0.00
6 T	Chloroethane	0.471	0.448	4.9	89	0.00
7 T	Trichlorofluoromethane	1.065	0.983	7.7	86	0.00
8 T	Diethyl Ether	0.281	0.269	4.3	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.601	0.566	5.8	86	-0.01
10 T	Methyl Iodide	0.616	0.536	13.0	72	0.00
11 T	Tert butyl alcohol	0.046	0.039	15.2	92	0.00
12 CM	1,1-Dichloroethene	0.549	0.516	6.0#	83	0.00
13 T	Acrolein	0.056	0.049	12.5	88	0.00
14 T	Allyl chloride	1.047	0.995	5.0	85	0.00
15 T	Acrylonitrile	0.143	0.142	0.7	87	0.00
16 T	Acetone	0.173	0.198	-14.5	96	0.00
17 T	Carbon Disulfide	1.915	1.724	10.0	83	0.00
18 T	Methyl Acetate	0.374	0.328	12.3	83	0.00
19 T	Methyl tert-butyl Ether	1.314	1.339	-1.9	88	0.00
20 T	Methylene Chloride	0.713	0.610	14.4	85	0.00
21 T	trans-1,2-Dichloroethene	0.639	0.612	4.2	86	0.00
22 T	Diisopropyl ether	2.052	2.134	-4.0	89	0.00
23 T	Vinyl Acetate	1.132	1.214	-7.2	88	0.00
24 P	1,1-Dichloroethane	1.196	1.162	2.8	86	0.00
25 T	2-Butanone	0.202	0.220	-8.9	92	0.00
26 T	2,2-Dichloropropane	0.990	0.960	3.0	87	0.00
27 T	cis-1,2-Dichloroethene	0.687	0.667	2.9	86	0.00
28 T	Bromochloromethane	0.451	0.442	2.0	86	0.00
29 T	Tetrahydrofuran	0.119	0.123	-3.4	88	0.00
30 C	Chloroform	1.222	1.194	2.3#	89	0.00
31 T	Cyclohexane	1.207	1.120	7.2	87	0.00
32 T	1,1,1-Trichloroethane	1.078	1.029	4.5	88	0.00
33 S	1,2-Dichloroethane-d4	0.632	0.655	-3.6	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.329	0.348	-5.8	98	0.00
36 T	1,1-Dichloropropene	0.564	0.554	1.8	87	0.00
37 T	Ethyl Acetate	0.262	0.258	1.5	86	0.00
38 T	Carbon Tetrachloride	0.558	0.544	2.5	87	0.00
39 T	Methylcyclohexane	0.654	0.646	1.2	86	0.00
40 TM	Benzene	1.523	1.502	1.4	87	0.00
41 T	Methacrylonitrile	0.148	0.165	-11.5	91	0.00
42 TM	1,2-Dichloroethane	0.481	0.472	1.9	86	0.00
43 T	Isopropyl Acetate	0.471	0.499	-5.9	90	0.00
44 TM	Trichloroethene	0.393	0.378	3.8	87	0.00
45 C	1,2-Dichloropropane	0.393	0.394	-0.3#	89	0.00
46 T	Dibromomethane	0.206	0.208	-1.0	88	0.00
47 T	Bromodichloromethane	0.544	0.548	-0.7	89	0.00
48 T	Methyl methacrylate	0.249	0.249	0.0	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020222\
 Data File : VD071883.D
 Acq On : 02 Feb 2022 17:02
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 01:08:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	90	0.00
50 S	Toluene-d8	1.227	1.297	-5.7	95	0.00
51 T	4-Methyl-2-Pentanone	0.247	0.263	-6.5	90	0.00
52 CM	Toluene	0.947	0.951	-0.4#	88	0.00
53 T	t-1,3-Dichloropropene	0.499	0.504	-1.0	86	0.00
54 T	cis-1,3-Dichloropropene	0.585	0.579	1.0	87	0.00
55 T	1,1,2-Trichloroethane	0.273	0.275	-0.7	88	0.00
56 T	Ethyl methacrylate	0.338	0.357	-5.6	87	0.00
57 T	1,3-Dichloropropane	0.487	0.499	-2.5	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.155	2.5	82	0.00
59 T	2-Hexanone	0.179	0.202	-12.8	92	0.00
60 T	Dibromochloromethane	0.337	0.342	-1.5	88	0.00
61 T	1,2-Dibromoethane	0.263	0.268	-1.9	88	0.00
62 S	4-Bromofluorobenzene	0.429	0.464	-8.2	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.354	0.335	5.4	87	0.00
65 PM	Chlorobenzene	1.058	1.043	1.4	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.378	1.8	89	0.00
67 C	Ethyl Benzene	1.946	1.963	-0.9#	89	0.00
68 T	m/p-Xylenes	0.740	0.752	-1.6	89	0.00
69 T	o-Xylene	0.673	0.680	-1.0	88	0.00
70 T	Styrene	1.160	1.195	-3.0	90	0.00
71 P	Bromoform	0.205	0.204	0.5	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.872	3.741	3.4	88	0.00
74 T	N-amyl acetate	0.952	1.002	-5.3	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.716	0.687	4.1	92	0.00
76 T	1,2,3-Trichloropropane	0.527	0.536	-1.7	100	0.00
77 T	Bromobenzene	0.861	0.827	3.9	90	0.00
78 T	n-propylbenzene	4.855	4.786	1.4	90	0.00
79 T	2-Chlorotoluene	2.751	2.681	2.5	90	0.00
80 T	1,3,5-Trimethylbenzene	3.211	3.179	1.0	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.214	0.200	6.5	86	0.00
82 T	4-Chlorotoluene	2.862	2.775	3.0	90	0.00
83 T	tert-Butylbenzene	2.661	2.660	0.0	91	0.00
84 T	1,2,4-Trimethylbenzene	3.159	3.174	-0.5	90	0.00
85 T	sec-Butylbenzene	4.164	4.150	0.3	91	0.00
86 T	p-Isopropyltoluene	3.384	3.432	-1.4	91	0.00
87 T	1,3-Dichlorobenzene	1.749	1.694	3.1	91	0.00
88 T	1,4-Dichlorobenzene	1.745	1.677	3.9	91	0.00
89 T	n-Butylbenzene	3.320	3.355	-1.1	93	0.00
90 T	Hexachloroethane	0.673	0.666	1.0	94	0.00
91 T	1,2-Dichlorobenzene	1.497	1.457	2.7	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.119	0.117	1.7	93	0.00
93 T	1,2,4-Trichlorobenzene	0.874	0.856	2.1	91	0.00
94 T	Hexachlorobutadiene	0.477	0.468	1.9	95	0.00
95 T	Naphthalene	1.588	1.637	-3.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020222\
Data File : VD071883.D
Acq On : 02 Feb 2022 17:02
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Feb 03 01:08:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.751	-0.3	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6