


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071872\
Data File : VD071872.D
Acq On : 02 Feb 2022 08:59
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Title : SW846 8260
Last Update : Thu Jan 27 06:04:38 2022



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	23.8	70723	PASS
75	95	30	60	56.1	167061	PASS
95	95	100	100	100.0	297621	PASS
96	95	5	9	6.5	19437	PASS
173	174	0.00	2	1.6	3545	PASS
174	95	50	100	74.4	221419	PASS
175	174	5	9	7.3	16231	PASS
176	174	95	101	95.8	212032	PASS
177	176	5	9	6.2	13246	PASS