

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020224\
 Data File : VD078331.D
 Acq On : 02 Feb 2024 17:27
 Operator : RP/MD
 Sample : P1329-14
 Misc : 3.68G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0FR5

Quant Time: Feb 03 00:23:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Feb 03 00:20:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	148414	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	112581	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	25470	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	105204	24.985	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.920%
7) Chloroethane-d5	2.799	69	88584	25.027	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.120%
11) 1,1-Dichloroethene-d2	3.911	65	19579	21.535	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	86.160%
21) 2-Butanone-d5	6.987	46	20553	33.152	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	66.300%
24) Chloroform-d	7.569	84	94895	22.426	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.720%
26) 1,2-Dichloroethane-d4	8.222	65	52110	22.819	ug/L	-0.01
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.280%
32) Benzene-d6	8.199	84	187227	28.442	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	113.760%
36) 1,2-Dichloropropane-d6	9.210	67	60996	30.098	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	120.400%#
41) Toluene-d8	10.269	98	134838	22.805	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.200%
43) trans-1,3-Dichloroprop...	10.522	79	19688	21.757	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.040%
47) 2-Hexanone-d5	10.875	63	15291	37.188	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.380%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	42962	22.814	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.240%
66) 1,2-Dichlorobenzene-d4	13.810	152	17907	19.638	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020224\
Data File : VD078331.D
Acq On : 02 Feb 2024 17:27
Operator : RP/MD
Sample : P1329-14
Misc : 3.68G/10ml/MSVOA_D/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0FR5

Quant Time: Feb 03 00:23:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Feb 03 00:20:19 2024
Response via : Initial Calibration

