

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020224\
 Data File : VD078329.D
 Acq On : 02 Feb 2024 16:33
 Operator : RP/MD
 Sample : P1329-15
 Misc : 5.68G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0FR6

Quant Time: Feb 03 00:22:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Feb 03 00:20:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	152347	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	129565	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	38827	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	113802	26.329	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 105.320%	
7) Chloroethane-d5	2.799	69	96288	26.502	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 106.000%	
11) 1,1-Dichloroethene-d2	3.910	65	22399	24.001	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 96.000%	
21) 2-Butanone-d5	6.987	46	26864	42.213	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 84.420%	
24) Chloroform-d	7.563	84	110520	25.444	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 101.760%	
26) 1,2-Dichloroethane-d4	8.234	65	59767	25.497	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 102.000%	
32) Benzene-d6	8.198	84	220991	29.170	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 116.680%	
36) 1,2-Dichloropropane-d6	9.210	67	70737	30.329	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 121.320%#	
41) Toluene-d8	10.269	98	186168	27.359	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 109.440%	
43) trans-1,3-Dichloroprop...	10.528	79	25272	24.267	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 97.080%	
47) 2-Hexanone-d5	10.881	63	20939	44.249	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 88.500%	
56) 1,1,2,2-Tetrachloroeth...	12.651	84	55453	25.587	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 102.360%	
66) 1,2-Dichlorobenzene-d4	13.816	152	36468	26.236	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 104.960%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020224\
Data File : VD078329.D
Acq On : 02 Feb 2024 16:33
Operator : RP/MD
Sample : P1329-15
Misc : 5.68G/10ml/MSVOA_D/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0FR6

Quant Time: Feb 03 00:22:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Feb 03 00:20:19 2024
Response via : Initial Calibration

