

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020224\
 Data File : VD078334.D
 Acq On : 02 Feb 2024 18:50
 Operator : RP/MD
 Sample : P1329-20
 Misc : 5.83G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0FM5

Quant Time: Feb 03 00:23:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Feb 03 00:20:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	143767	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	141219	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	54329	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	106570	26.127	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.520%
7) Chloroethane-d5	2.799	69	89574	26.125	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.520%
11) 1,1-Dichloroethene-d2	3.916	65	21956	24.930	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	99.720%
21) 2-Butanone-d5	6.993	46	29286	48.765	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.540%
24) Chloroform-d	7.569	84	100612	24.545	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	8.234	65	58788	26.576	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.320%
32) Benzene-d6	8.199	84	204061	24.713	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.840%
36) 1,2-Dichloropropane-d6	9.210	67	63891	25.133	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.520%
41) Toluene-d8	10.269	98	177811	23.974	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.880%
43) trans-1,3-Dichloroprop...	10.528	79	27566	24.285	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.160%
47) 2-Hexanone-d5	10.881	63	24506	47.513	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.020%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	57657	24.408	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.640%
66) 1,2-Dichlorobenzene-d4	13.816	152	50134	25.776	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020224\
Data File : VD078334.D
Acq On : 02 Feb 2024 18:50
Operator : RP/MD
Sample : P1329-20
Misc : 5.83G/10ml/MSVOA_D/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0FM5

Quant Time: Feb 03 00:23:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM020124SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Feb 03 00:20:19 2024
Response via : Initial Calibration

