

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
 Data File : VD071912.D
 Acq On : 08 Feb 2022 12:29
 Operator : VA/SY
 Sample : N1423-02
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-21

Quant Time: Feb 09 03:49:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	23569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	43017	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	42849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	19462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	24440	82.041	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 164.080%#	
35) Dibromofluoromethane	7.908	113	15567	55.053	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.100%	
50) Toluene-d8	10.338	98	51522	48.789	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 97.580%	
62) 4-Bromofluorobenzene	12.620	95	18467	50.076	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
Data File : VD071912.D
Acq On : 08 Feb 2022 12:29
Operator : VA/SY
Sample : N1423-02
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-21

Quant Time: Feb 09 03:49:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

