

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
 Data File : VD071915.D
 Acq On : 08 Feb 2022 13:55
 Operator : VA/SY
 Sample : N1423-08
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-24

Quant Time: Feb 09 03:50:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	326461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	626287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	612706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	270846	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	251231	60.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.780%
35) Dibromofluoromethane	7.914	113	219016	53.201	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.400%
50) Toluene-d8	10.338	98	799522	52.002	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.000%
62) 4-Bromofluorobenzene	12.626	95	264478	49.260	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
Data File : VD071915.D
Acq On : 08 Feb 2022 13:55
Operator : VA/SY
Sample : N1423-08
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-24

Quant Time: Feb 09 03:50:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

