

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
 Data File : VD071917.D
 Acq On : 08 Feb 2022 14:52
 Operator : VA/SY
 Sample : N1424-02
 Misc : 6.29G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DB-10(22-25)-02-04-22

Quant Time: Feb 09 03:51:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	438683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	834764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	792789	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	349941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	294679	53.146	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.300%	
35) Dibromofluoromethane	7.909	113	280280	51.079	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.160%	
50) Toluene-d8	10.332	98	1068163	52.124	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.240%	
62) 4-Bromofluorobenzene	12.626	95	360902	50.431	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.860%	
Target Compounds						Qvalue
16) Acetone	4.156	43	51691	34.056	ug/l	91
17) Carbon Disulfide	4.409	76	41947	2.496	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
Data File : VD071917.D
Acq On : 08 Feb 2022 14:52
Operator : VA/SY
Sample : N1424-02
Misc : 6.29G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DB-10(22-25)-02-04-22

Quant Time: Feb 09 03:51:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

