

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
 Data File : VD071918.D
 Acq On : 08 Feb 2022 15:20
 Operator : VA/SY
 Sample : N1424-03
 Misc : 6.75G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BLIND-DUPLICATE

Quant Time: Feb 09 03:51:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	422027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	799802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	755475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	338685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	268778	50.388	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.780%	
35) Dibromofluoromethane	7.914	113	267591	50.899	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.800%	
50) Toluene-d8	10.332	98	1031814	52.551	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.100%	
62) 4-Bromofluorobenzene	12.626	95	342130	49.898	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 99.800%	
Target Compounds						
						Qvalue
16) Acetone	4.156	43	23371	16.005	ug/l	# 89
17) Carbon Disulfide	4.403	76	34924	2.160	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
Data File : VD071918.D
Acq On : 08 Feb 2022 15:20
Operator : VA/SY
Sample : N1424-03
Misc : 6.75G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BLIND-DUPLICATE

Quant Time: Feb 09 03:51:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

