

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
 Data File : VD071924.D
 Acq On : 08 Feb 2022 18:12
 Operator : VA/SY
 Sample : N1423-10RE
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-25RE

Quant Time: Feb 09 04:01:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	5424	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	8536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	8185	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	3735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	5250	76.579	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 153.160%	
35) Dibromofluoromethane	7.914	113	2722	48.512	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.020%	
50) Toluene-d8	10.338	98	10744	51.272	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.540%	
62) 4-Bromofluorobenzene	12.626	95	3745	51.177	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020822\
Data File : VD071924.D
Acq On : 08 Feb 2022 18:12
Operator : VA/SY
Sample : N1423-10RE
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-25RE

Quant Time: Feb 09 04:01:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

