

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020824\
 Data File : VD078379.D
 Acq On : 08 Feb 2024 19:50
 Operator : RP/MD
 Sample : P1394-05
 Misc : 4.75G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DUP-020624

Quant Time: Feb 09 03:14:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D013024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 03:58:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	57665	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	134506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	157305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	74599	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	45938	64.352	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.700%
35) Dibromofluoromethane	7.804	113	50086	52.304	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.600%
50) Toluene-d8	10.269	98	178392	49.205	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.420%
62) 4-Bromofluorobenzene	12.569	95	65028	59.894	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	119.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020824\
Data File : VD078379.D
Acq On : 08 Feb 2024 19:50
Operator : RP/MD
Sample : P1394-05
Misc : 4.75G/5ml/MSVOA_D/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DUP-020624

Quant Time: Feb 09 03:14:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D013024S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 03:58:19 2024
Response via : Initial Calibration

