

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020924\
 Data File : VD078388.D
 Acq On : 09 Feb 2024 16:31
 Operator : RP/MD
 Sample : P1394-03RE
 Misc : 4.77G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3(4-5)RE

Quant Time: Feb 09 23:02:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D013024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 03:58:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	95463	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	219846	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	240744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	111127	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	57749	48.867	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.740%
35) Dibromofluoromethane	7.804	113	70572	45.090	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.180%
50) Toluene-d8	10.269	98	291887	49.258	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.569	95	97777	55.099	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD020924\
Data File : VD078388.D
Acq On : 09 Feb 2024 16:31
Operator : RP/MD
Sample : P1394-03RE
Misc : 4.77G/5ml/MSVOA_D/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3(4-5)RE

Quant Time: Feb 09 23:02:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D013024S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 03:58:19 2024
Response via : Initial Calibration

