



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD078382\
Data File : VD078382.D
Acq On : 09 Feb 2024 09:10
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D0130245.M
Title : SW846 8260
Last Update : Thu Feb 01 03:58:19 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.7	14355	PASS
75	95	30	60	53.8	39232	PASS
95	95	100	100	100.0	72864	PASS
96	95	5	9	6.3	4626	PASS
173	174	0.00	2	1.6	982	PASS
174	95	50	100	83.1	60523	PASS
175	174	5	9	7.3	4404	PASS
176	174	95	101	95.9	58021	PASS
177	176	5	9	6.7	3879	PASS