

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021422\
 Data File : VD071954.D
 Acq On : 14 Feb 2022 13:54
 Operator : VA/SY
 Sample : N1521-01
 Misc : 6.64G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP1-A

Quant Time: Feb 15 00:35:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	395159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	764024	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	734056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	313085	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	258392	51.734	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.460%
35) Dibromofluoromethane	7.914	113	252643	50.306	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.620%
50) Toluene-d8	10.338	98	995284	53.065	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.120%
62) 4-Bromofluorobenzene	12.626	95	321518	49.088	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021422\
Data File : VD071954.D
Acq On : 14 Feb 2022 13:54
Operator : VA/SY
Sample : N1521-01
Misc : 6.64G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP1-A

Quant Time: Feb 15 00:35:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

