

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021422\
 Data File : VD071964.D
 Acq On : 14 Feb 2022 18:33
 Operator : VA/SY
 Sample : N1525-01RE
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-021122RE

Quant Time: Feb 15 00:38:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	82199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	153755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	168743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	76749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	90814	87.409	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 174.820%#	
35) Dibromofluoromethane	7.914	113	63190	62.523	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 125.040%	
50) Toluene-d8	10.338	98	205838	54.533	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.060%	
62) 4-Bromofluorobenzene	12.626	95	72164	54.748	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.500%	
Target Compounds						Qvalue
16) Acetone	4.167	43	10368	36.455	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021422\
Data File : VD071964.D
Acq On : 14 Feb 2022 18:33
Operator : VA/SY
Sample : N1525-01RE
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-021122RE

Quant Time: Feb 15 00:38:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

