

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021722\
 Data File : VD072012.D
 Acq On : 17 Feb 2022 14:32
 Operator : VA/SY
 Sample : N1579-04
 Misc : 5.52G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-6-(6.5-7.0)

Quant Time: Feb 18 06:37:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	240320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	468447	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	425875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	169979	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	169341	55.750	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.500%
35) Dibromofluoromethane	7.908	113	152906	49.657	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.320%
50) Toluene-d8	10.332	98	614053	53.396	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.800%
62) 4-Bromofluorobenzene	12.620	95	201101	50.076	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021722\
Data File : VD072012.D
Acq On : 17 Feb 2022 14:32
Operator : VA/SY
Sample : N1579-04
Misc : 5.52G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-6-(6.5-7.0)

Quant Time: Feb 18 06:37:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

