



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022024\
Data File : VD078472.D
Acq On : 20 Feb 2024 08:51
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
Title : SW846 8260
Last Update : Wed Feb 14 01:16:20 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	10105	PASS
75	95	30	60	52.1	26851	PASS
95	95	100	100	100.0	51525	PASS
96	95	5	9	7.2	3732	PASS
173	174	0.00	2	1.9	844	PASS
174	95	50	100	88.3	45509	PASS
175	174	5	9	8.0	3621	PASS
176	174	95	101	96.9	44115	PASS
177	176	5	9	6.0	2627	PASS