

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072030.D
 Acq On : 21 Feb 2022 14:08
 Operator : VA/SY
 Sample : N1610-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-021822

Quant Time: Feb 22 00:35:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/22/2022
 Supervised By :Mahesh Dadoda 02/23/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.973	168	609120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1050911	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	985539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	452092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	305540	48.357	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.720%
35) Dibromofluoromethane	7.909	113	318712	50.056	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.120%
50) Toluene-d8	10.332	98	1271638	53.855	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.720%
62) 4-Bromofluorobenzene	12.626	95	452345	52.153	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.300%
Target Compounds						
11) Tert butyl alcohol	5.197	59	10210m	28.321	ug/l	
17) Carbon Disulfide	4.409	76	5621	1.019	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072030.D
 Acq On : 21 Feb 2022 14:08
 Operator : VA/SY
 Sample : N1610-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :

MSVOA_D

ClientSampleId :

TR-05-021822

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/22/2022

Supervised By :Mahesh Dadoda 02/23/2022

Quant Time: Feb 22 00:35:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

