

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072031.D
 Acq On : 21 Feb 2022 14:36
 Operator : VA/SY
 Sample : N1626-02
 Misc : 4.47G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-B1

Quant Time: Feb 22 00:37:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	40594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	70472	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	68380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	28372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	34048	80.859	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	161.720%
35) Dibromofluoromethane	7.909	113	14722	34.480	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	68.960%
50) Toluene-d8	10.338	98	80939	51.118	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.240%
62) 4-Bromofluorobenzene	12.626	95	28952	49.778	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.560%
Target Compounds						
						Qvalue
16) Acetone	4.150	43	24877	273.365	ug/l	96
25) 2-Butanone	7.214	43	7623	60.607	ug/l #	88
59) 2-Hexanone	10.979	43	3310	16.395	ug/l	93
84) 1,2,4-Trimethylbenzene	13.261	105	3493	2.184	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072031.D
Acq On : 21 Feb 2022 14:36
Operator : VA/SY
Sample : N1626-02
Misc : 4.47G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-B1

Quant Time: Feb 22 00:37:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

