

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072033.D
 Acq On : 21 Feb 2022 15:33
 Operator : VA/SY
 Sample : N1626-04
 Misc : 4.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-B2

Quant Time: Feb 22 01:05:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	16096	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	26862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	27392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	11223	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	13060	78.221	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 156.440%	
35) Dibromofluoromethane	7.908	113	7471	45.905	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 91.820%	
50) Toluene-d8	10.332	98	31754	52.613	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.220%	
62) 4-Bromofluorobenzene	12.626	95	11117	50.145	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.280%	
Target Compounds						
16) Acetone	4.156	43	4559	126.345	ug/l	# 66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072033.D
Acq On : 21 Feb 2022 15:33
Operator : VA/SY
Sample : N1626-04
Misc : 4.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-B2

Quant Time: Feb 22 01:05:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

