

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
 Data File : VD072034.D
 Acq On : 21 Feb 2022 16:01
 Operator : VA/SY
 Sample : N1626-05
 Misc : 6.15G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CC-A3

Quant Time: Feb 22 01:05:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	534868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	931007	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	877002	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	400363	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	278049	50.12	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.24%	
35) Dibromofluoromethane	7.914	113	147595	26.17	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 52.34%#	
50) Toluene-d8	10.338	98	1137410	54.37	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.74%	
62) 4-Bromofluorobenzene	12.626	95	413604	53.83	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.66%	
Target Compounds						
						Qvalue
16) Acetone	4.155	43	19970	16.65	ug/l	96
93) 1,2,4-Trichlorobenzene	15.149	180	9318	1.33	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022122\
Data File : VD072034.D
Acq On : 21 Feb 2022 16:01
Operator : VA/SY
Sample : N1626-05
Misc : 6.15G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CC-A3

Quant Time: Feb 22 01:05:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

